


HPLC REPORT

Sample: K-14P
Analyst: WJ
Lot. No.: P140705-CJ594041
Column: Gemini-NX 5 μ C18 110A, 4.6*250mm
Solvent A: A: 0.1% Trifluoroacetic Acid in 100% Acetonitrile
Solvent B: B: 0.1% Trifluoroacetic Acid in 100% Water
Gradient:

Analyzed date: 2014-7-14

	A	B
0.0min	10%	90%
25.0min	70%	30%
25.1min	100%	0%
30.0min	Stop	

Volume: 20 μ l
Wavelength: 220nm
Flow rate: 1.0ml/min
mV



Rank	Time	Conc.	Area	Height
1	10.382	1.9629	97080	10190
2	11.190	2.2488	111218	24058
3	11.362	91.9667	4548424	434066
4	11.923	2.4134	119359	11020
5	12.116	1.4082	69648	6481
Total		100	4945729	485815